


AD-A127 997 NMR (NUCLEAR MAGNETIC RESONANCE) SENSOR FOR CALIBRATING 1/1  
THE FIELD OF A SU... (U) FOREIGN TECHNOLOGY DIV  
WRIGHT-PATTERSON AFB OH V B PLUZHNIKOV 04 MAR 83  
UNCLASSIFIED FTD-ID(RS)T-0104-83 F/G 9/1 NL



END  
DATE  
FILMED  
DTIC



MICROCOPY RESOLUTION TEST CHART  
NATIONAL BUREAU OF STANDARDS-1963-A

DA 127997


# FOREIGN TECHNOLOGY DIVISION



NMR SENSOR FOR CALIBRATING THE FIELD OF A  
SUPERCONDUCTING SOLENOID

by

V.B. Pluzhnikov



DTIC  
ELECTE

MAY 11 1983

A

Approved for public release;  
distribution unlimited.

DTIC FILE COPY

83 05 09-034

# EDITED TRANSLATION

FTD-ID(RS)T-0104-83

4 March 1983

MICROFICHE NR: FTD-83-C-000331

NMR SENSOR FOR CALIBRATING THE FIELD OF A  
SUPERCONDUCTING SOLENOID

By: V.B. Pluzhnikov

English pages: 3

Source: Sbornik Nauchnykh Trudov Fiziko-  
Tekhnicheskii Institut Nizkikh  
Temperatur, Akademiya Nauk USSR,  
Vol. 18, Kar'kov, 1972, pp. 15-16

Country of origin: USSR

Translated by: Charles T. Ostertag, Jr.

Requester: FTD/TQTD

Approved for public release; distribution unlimited.

Accession For	
DTIC	<input checked="" type="checkbox"/>
CRA&I	<input type="checkbox"/>
AS T&S	<input type="checkbox"/>
Unprocessed	<input type="checkbox"/>
Classification	
Availability Codes	
Avail and/or	
Special	

*A*



THIS TRANSLATION IS A RENDITION OF THE ORIGINAL FOREIGN TEXT WITHOUT ANY ANALYTICAL OR EDITORIAL COMMENT. STATEMENTS OR THEORIES ADVOCATED OR IMPLIED ARE THOSE OF THE SOURCE AND DO NOT NECESSARILY REFLECT THE POSITION OR OPINION OF THE FOREIGN TECHNOLOGY DIVISION.

PREPARED BY:

TRANSLATION DIVISION  
FOREIGN TECHNOLOGY DIVISION  
WP-AFB, OHIO.

FTD-ID(RS)T-0104-83

Date 4 Mar 19 83

# U. S. BOARD ON GEOGRAPHIC NAMES transliteration SYSTEM

Block	Italic	Transliteration	Block	Italic	Transliteration
А а	<i>А а</i>	A, a	Р р	<i>Р р</i>	R, r
Б б	<i>Б б</i>	B, b	С с	<i>С с</i>	S, s
В в	<i>В в</i>	V, v	Т т	<i>Т т</i>	T, t
Г г	<i>Г г</i>	G, g	У у	<i>У у</i>	U, u
Д д	<i>Д д</i>	D, d	Ф ф	<i>Ф ф</i>	F, f
Е е	<i>Е е</i>	Ye, ye; E, e*	Х х	<i>Х х</i>	Kh, kh
Ж ж	<i>Ж ж</i>	Zh, zh	Ц ц	<i>Ц ц</i>	Ts, ts
З з	<i>З з</i>	Z, z	Ч ч	<i>Ч ч</i>	Ch, ch
И и	<i>И и</i>	I, i	Ш ш	<i>Ш ш</i>	Sh, sh
Й й	<i>Й й</i>	Y, y	Щ щ	<i>Щ щ</i>	Shch, shch
К к	<i>К к</i>	K, k	Ъ ъ	<i>Ъ ъ</i>	"
Л л	<i>Л л</i>	L, l	Ы ы	<i>Ы ы</i>	Y, y
М м	<i>М м</i>	M, m	Ь ь	<i>Ь ь</i>	'
Н н	<i>Н н</i>	N, n	Э э	<i>Э э</i>	E, e
О о	<i>О о</i>	O, o	Ю ю	<i>Ю ю</i>	Yu, yu
П п	<i>П п</i>	P, p	Я я	<i>Я я</i>	Ya, ya

\*ye initially, after vowels, and after Ъ, ь; e elsewhere.  
When written as ё in Russian, transliterate as yě or ě.

## RUSSIAN AND ENGLISH TRIGONOMETRIC FUNCTIONS

Russian	English	Russian	English	Russian	English
sin	sin	sh	sinh	arc sh	sinh <sup>-1</sup>
cos	cos	ch	cosh	arc ch	cosh <sup>-1</sup>
tg	tan	th	tanh	arc th	tanh <sup>-1</sup>
ctg	cot	cth	coth	arc cth	coth <sup>-1</sup>
sec	sec	sch	sech	arc sch	sech <sup>-1</sup>
cosec	csc	csch	csch	arc csch	csch <sup>-1</sup>

Russian English

rot curl  
lg log

### GRAPHICS DISCLAIMER

All figures, graphics, tables, equations, etc. merged into this translation were extracted from the best quality copy available.

## NMR SENSOR FOR CALIBRATING THE FIELD OF A SUPERCONDUCTING SOLENOID

V. B. Pluzhnikov

Submitted 28 April 1972

A description is given of a simple construction of a magnetic-field sensor, making it possible to carry out the absolute calibration of the field of a superconducting solenoid by the NMR method with the help of the IMI-2 (Ye11-2) device which is produced by or own [Soviet] industry.

In recent years abroad the method of nuclear magnetic resonance [NMR] has received wide application for the absolute calibration of the field of superconducting solenoids [1-3]; in our country this has been hindered by the absence of a series-produced gaussmeter with a sensor for operation in liquid helium.

A description is given below of the construction of a NMR sensor for the series-produced [Soviet] IMI-2 (Ye11-2) gaussmeter, used by us for the calibration of the field of a superconducting solenoid in experiments on studying the De Haas-Van Alphen effect (DKhVA) on transition metals [4].

On a paper housing with a diameter and length of 3 mm each a coil made out of nine windings of PELShO brand wire 0.1 mm in diameter was wound, and inside paraffin with a large amount of fine aluminum powder was poured. The dimensions of the grains of powder should be of the order of depth of the skin layer for obtaining the maximum NMR signal from that particular volume of the sample. It turned out

that the commercial pigment "serebryanka" [silver steel] is completely suitable for this purpose. The sample prepared in this manner was inserted in a fluoroplastic holder, secured on the end of a rigid coaxial cable, connecting the coil mentioned above with the input of a gaussmeter autodyne, and thus forming its high-frequency oscillatory circuit. The low-temperature section of the cable 1.2 m long was fabricated out of two thin-walled tubes of stainless steel 16 mm and 3 mm in diameter (for reducing heat input and the natural capacitance of the cable), and outside of the cryostat this was ordinary coaxial cable 0.3 m long. The range of frequencies spanned by the autodyne with this NMR sensor turned out to be 5.7-17.4 megahertz and was measured with a Ch4-1 heterodyne wavemeter.

The small modulating field of  $\sim 30$  G, necessary for observing the NMR signal, was created by a cylindrical coil 30 mm in length, powered from a network through a LATR [laboratory autotransformer] and a filament transformer. The axis of the modulating coil coincided with the direction of the magnetic field in the solenoid and was perpendicular to the axis of the high-frequency coil around the sample.

A resonance curve was observed on the screen of the S1-19 oscillograph in fields of 2-4 kG from protons ( $H^1$ ) in paraffin and from the nuclei of fluoron in fluoroplast (material of the sample holder), and further at 5-16 kG from nuclei of  $Al^{27}$  [3]. Signal/noise ratio = 10 V, field 12 kG. Using the point of intersection of the two branches of the resonance curve for marking the field, it is easy to distinguish a change in the field of 1 G. The device was used for calibrating the field of an uncorrected solenoid of the cylindrical type with geometric dimensions of the winding: length - 156 mm, outer diameter - 115 mm, inner diameter - 25 mm, made with a seven-core cable 1 mm in diameter which was made of the superconducting alloy 65-BT. The housing of the solenoid was fabricated from nonmagnetic stainless steel. The current in the winding was measured with the help of a PP-63 potentiometer and an F116 device. The accuracy of recording the current produced an error in the determination of the field of  $\pm 10$  G. During calibration of the field in the solenoid it turned out that its constant  $C=H$ , kG/I, and depends on the field in the same manner as in [1], and its value obtained at 16 kG differed

by 0.2% from its magnitude in large fields (greater than 30 kG), determined based on the DKhVA effect [5]. Determination of hysteresis in the solenoid showed that the maximum residual field (on the level of 2 kG) does not exceed 60 G; for solenoids with a Nb-Zr winding it is considerably greater [6].

In conclusion let us note that the method described for the absolute calibration of the field of superconducting solenoids is convenient and does not require special experience.

I express my sincere thanks to I. V. Svechkarev for attention to the work and to V. V. Zhukov and P. S. Kalinin for useful information on the experimental method.

#### REFERENCES

1. Максфилд и Меррил. Приборы для научных исследований, № 8, 5, 1965.
2. Рэпп мл., Приборы для научных исследований, № 8, 52, 1966.
3. Хиггинс и Чанг, Приборы для научных исследований, № 4, 82, 1968.
4. V.B.Pluzhnikov and I.V.Svechkarev, Phys.Stat.Sol., (b), 43, K55, 1971.
5. O'Sullivan and W.J.Schirber, Cryogenics, 7, 118, 1967.
6. Алдерсон и Саренинский, Приборы для научных исследований, № 3, 82, 1963.

# DISTRIBUTION LIST

## DISTRIBUTION DIRECT TO RECIPIENT

<u>ORGANIZATION</u>	<u>MICROFICHE</u>
A205 DMAHTC	1
A210 DMAAC	1
B344 DIA/RTS-2C	9
C043 USAMIIA	1
C500 TRADOC	1
C509 BALLISTIC RES LAB	1
C510 R&T LABS/AVRADCOM	1
C513 ARRADCOM	1
C535 AVRADCOM/TSARCOM	1
C539 TRASANA	1
C591 PSTC	4
C619 MIA REDSTONE	1
D008 NISC	1
E053 HQ USAF/INET	1
E403 AFSC/INA	1
E404 AEDC/DOF	1
E408 AFWL	1
E410 AD/IND	1
E429 SD/IND	1
P005 DOE/ISA/DDI	1
P050 CIA/OCR/ADD/SD	2
AFIT/LDE	1
FTD	
CCN	1
NIA/PHS	1
NIIS	2
LLNL/Code L-389	1
NASA/NST-44	1
NSA/1213/TDL	2